

Technical Specifications Curatron	2000PC
Pad 50 x 70 cm = 20" x 28"	option
High energy coil. Diameter 20 x 20 cm = 8"	option
Very high energy coil. Diameter 20 x 20 cm = 8"	option
Mattress 65 x 180 cm = 26" x 71"	option
Max energy Pad	40 milli-Tesla (400 Gauss)
Max energy high energy coil	70 milli-Tesla (700 Gauss)
Max energy very high energy coil	100 milli-Tesla (1000 Gauss)
Max energy mattress	15 milli-Tesla (150 Gauss)
Master programs	10 + 70
Program time for 10 programs	20 – 30 min.
Frequency range	1 – 50 Hz
User programmable master programs	Yes
Patient database and Therapy protocols databases	Yes
Minimum stage time for each parameter setting	1 min.
Maximum stage time for each parameter setting	20 min.
Minimum total therapy time per session	1 min.
Maximum total therapy time per session	99 min